

AN1004 – Extraction of Log Data from Accu-Gold



- Start Accu-Gold 2
- In the "Accu-Gold 2 Start Menu" click on "Continue":

Accu-Gold 2	Start		-		×
	8	Connect via Wi-Fi Connect via USB Open Session]		
Radcal [®] Worldwide leader in Diagnostic X-Ray Measurement www.radcaLcom		Cont	inue	→	

• Click on the "Accu-Gold 2" menu in the upper left corner:

\equiv ACCU-GOL	.D 2	New		_ = ×	
		st Sensors			
	Menu 🗲	ment	(\mathbf{F}))	
Main					
New	Ctrl+N	Duration	Cor	nment ⊑• ြ ြ	
Open	Ctrl+O		Ad Ad	d a Note	
Import		Rate AGMS			
			<u> </u>		
Save as					
Licenses		Pulse Count			
Help					
Close App					
			\$	Carry over to new measurements	
Accu-Gold not connected Image: Connected for the conne					



AN1004 – Extraction of Log Data from Accu-Gold



- Click on the menu entry "Help"
- Click on the menu entry "About"

	×
Accu-Gold 2	
Version 2.32 Build 4	
 Accu-Gold is up to date 	\mathbf{x}
Download Accu-Gold from radcal.com	
Visit www.radcal.com	
Send feedback	
Export log	
Copyright © 2016 Radcal. All rights reserved. Logima Software is Radcal's development partner.	

- Click on "Export log..."
- Save the "Accu-Gold_Logs.zip" where you can find it again then email to Cust_sup@Radcal.com.