



Declaration of Conformity

According to Engineering Report # ER1190829-PDC

Manufacturer's Name: Radcal
Manufacturer's Address: 426 West Duarte Road
Monrovia, CA 91016
U.S.A.

Declares that the Product
Product Name: Patient Dose Calibrator
Model Number(s): PDC-01
Product Accessories: All

Conforms to the following Product Specification:
EMC: *

EN 61326-1 (2013) Group 1 Class B
EN 61000-4-2
EN 61000-4-3
EN 61000-4-4
EN 61000-4-6
EN 61000-4-5
EN 61000-4-11
EN 61000-3-2
EN 61000-3-3

PERFORMANCE:
IEC 61674

The product herewith complies with the RFI-emissions requirements and immunity requirements and carries the CE marking accordingly.

E. MacIntosh
Technical Manager
August 29, 2019

* Tested with 3m USB cable